


CERTIFICATE OF CALIBRATION

ISSUED BY **Cirrus Research plc**

DATE OF ISSUE **13 February 2024**

CERTIFICATE NUMBER **208440**



Cirrus Research plc
Acoustic House
Bridlington Road
Hunmanby
North Yorkshire
YO14 0PH
United Kingdom

Page 1 of 2

Approved signatory

M.berry

Electronically signed:

M.BERRY

Sound Level Meter : IEC 61672-3:2013

Instrument information

Manufacturer: Cirrus Research plc

Notes:

Model: CR:900

Serial number: QT000113

Class: 2

Firmware version: 5.6.3196

Test summary

Date of calibration: 06 February 2024

The calibration was performed respecting the requirements of ISO/IEC 17025:2017.

Periodic tests were performed in accordance with procedures from IEC 61672-3:2013.

The sound level meter submitted for testing successfully completed the class 2 periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed.

However, no general statement or conclusion can be made about conformance of the sound level meter to the full specifications of IEC 61672-1:2013 because (a) evidence was not publicly available, from an independent testing organisation responsible for pattern approvals, to determine that the model of sound level meter fully conformed to the class 2 specifications in IEC 61672-1:2013 or correction data for acoustical test of frequency weighting were not provided in the Instruction Manual and (b) because the periodic tests of IEC 61672-3:2013 cover only a limited subset of the specifications in IEC 61672-1:2013.

Notes

This certificate provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. The results within this certificate relate only to the items calibrated. The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%.

CERTIFICATE OF CALIBRATION

Certificate Number:

208440

Page 2 of 2

Environmental conditions

The following conditions were recorded at the time of the test:

Before Pressure: 99.80 kPa Temperature: 21.0 °C Humidity: 45 %
After Pressure: 99.70 kPa Temperature: 21.0 °C Humidity: 44 %

Test equipment

Equipment	Manufacturer	Model	Serial number
Signal Generator	TTi	TG4001	414138
Attenuator	Cirrus Research	ZE:952	78711

Additional instrument information

Instruction manual:

Reference level range: Single range

Pattern approval: No

Source of pattern approval: -

Preamplifier

Model: MV:200F

Serial number: 9493F

Microphone

Model: MK:224

Serial number: 218024A

Test results summary

Test	Result
------	--------