



Comparison Report

The instrument detailed below has been tested by a Cirrus Engineer using independent equipment

Customer: M.Sport

Instrument type under test

Gill 1405-PK-021 Wind Monitor

Instrument Serial Number: 19390076

Tested by:

M.Berry

Signed

A handwritten signature in blue ink that reads 'M. BERRY' with a horizontal line underneath.

Test date:

22/03/2022

Test Procedure

Details of test equipment & equipment under test are shown below including results, shown below.

Model	Engineer Compass Directional Compass 12x	Gill 1405-PK-021 (W/D)
Serial No.	NSNA	19390076
Cal Cert No.	N/A	N/A
Wind Direction	0°	0°
Model	Thermo- Anemometer DT- 618B	Gill 1405-PK-021 (W/S)
Serial No.	11040620	19390076
Cal Cert No.	HAM48181	N/A
Wind Speed	2.09 m/s	2.20 m/s

