



# Comparison Report

The instrument detailed below has been tested by a Cirrus Engineer using independent equipment

**Customer: M.Sport**

**Instrument type under test**  
Gill 1405-PK-021 Wind Monitor

**Instrument Serial Number**  
19390076

Tested by:

M.Berry

Signed

A handwritten signature in blue ink that reads 'M. BERRY' with a horizontal line underneath.

Test date: 08/03/2023

**Test Procedure**

Details of test equipment & equipment under test are shown below including results, shown below.

<b>Model</b>	<b>Engineer Compass Directional Compass 12x</b>	<b>Gill 1405-PK-021 (W/D)</b>
<b>Serial No.</b>	NSNA	19390076
<b>Cal Cert No.</b>	N/A	N/A
<b>Wind Direction</b>	Due North	Due North
<b>Model</b>	<b>Thermo- Anemometer DT- 618B</b>	<b>Gill 1405-PK-021 (W/S)</b>
<b>Serial No.</b>	11040620	19390076
<b>Cal Cert No.</b>	(CLG) HAM55969	N/A
<b>Wind Speed</b>	4.04m/s	4.68 m/s

